



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2017-03-27**

Description: **SBE-4 Conductivity Sensor**

Sensor Information:

Replaced the main piston "O"-Rings.
Replaced conductivity cell.

Model Number: **SBE-4**

Serial Number: **04-4003**

Pressure Test Protocol:

Low Pressure Test: **50** PSI Held For: **15** Minutes

High Pressure Test: **10000** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **BLT**

**High pressure is
generally equal
to the maximum
depth rating of
the instrument**

