



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2017-03-07**

Description: **SBE-9plus CTD**

Sensor Information:

Replaced the main piston "O"-Rings.

Replaced 4-pin connector JT4.

Model Number: **SBE-9P**

Serial Number: **1087**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **10000** PSI Held For: **45** Minutes

Passed Test: **True**

Tested By: **TN**

High pressure is generally equal to the maximum depth rating of the instrument

